



PRESS RELEASE

2024-11-01

CONTINUOUS MICROBIAL MONITORING IN GRADE A CLEANROOM

The new official documentation related to the microbial monitoring in Grade A of Cleanroom requests the adoption of continuous monitoring. The adoption of the air sampler with 2 or 3 aspirating head is the solution for working shift of 3 or 6 or 9 hours. In fact:

- A. The intervention of operator with the air sampler is limited only to the initial preparation and final part reducing the contamination risk.
- B. There is the possibility to identify at which period a possible contamination happened
- C. The reduction of number of culture plates produces a simple and easy manipulation for the busy operator and a consistent saving cost
- D. The command unit is outside from the isolator for a more convenient operations
- E. There is also the possibility to have the culture plate outside from the isolator to simplify the activity of operator.

The stainless steel RABS with 2 or 3 satellites and command unit covers these basic needs.



TRIO.BAS RABS ISOLATOR CM + 3
SATELLITES CM